
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001296
8:39 01/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:39
STERILIZING 4:39
136.4 C 221kPa 4:39
136.1 C 219kPa 5:09
136.4 C 221kPa 5:39
136.4 C 223kPa 6:09
136.2 C 223kPa 6:39
136.4 C 224kPa 7:09
136.4 C 224kPa 7:39
136.4 C 221kPa 8:09
136.4 C 221kPa 8:09
VENTING 8:09
AIR DRYING 9:05
CYCLE COMPLETE 9:07

Digital Signature #
0E51378503EB1E51
