
Midmark M11 - v1.0.5
Total Cycles: 3904

Name: RITTER2
Sterilizer ID: V2313228
09/03/2024 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 86.7 F 0.0
4:00 101.3 F 0.0
6:00 138.6 F 0.4
8:00 165.0 F 1.6
10:00 181.9 F 4.6
12:00 196.8 F 7.8
14:00 208.6 F 11.2
16:00 216.4 F 6.0
18:00 226.2 F 8.3
20:00 236.3 F 12.5
22:00 235.8 F 9.4
24:00 240.0 F 10.3
26:00 246.9 F 13.9
28:00 258.2 F 20.1
30:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.1 F 28.8
2:00 272.6 F 29.1
3:00 272.7 F 29.5
4:00 271.8 F 29.1

Min 271.7 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 209.0 F 0.0
10:00 203.4 F 0.0
15:00 199.8 F 0.0
20:00 207.0 F 0.0
25:00 210.4 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 31:09
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:05
CYCLE COMPLETE

09/03/2024 12:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____