
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001940
14:28 09/08/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 1:34
PRESSURIZING 3:41
STERILIZING 4:28
136.0 C 220kPa 4:28
136.2 C 223kPa 4:58
136.2 C 222kPa 5:28
136.1 C 223kPa 5:58
136.0 C 220kPa 6:28
136.0 C 222kPa 6:58
136.4 C 225kPa 7:28
136.4 C 224kPa 7:58
136.3 C 224kPa 8:28
136.4 C 225kPa 8:58
136.4 C 225kPa 9:28
136.4 C 226kPa 9:58
136.4 C 226kPa 10:28
CYCLE ABORTED 10:52
STOP BUTTON PRESSED
NOT STERILE
END TIME 10:52

Digital Signature #
D85BECE831D5110B
