

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4365

Name: RITTER 1  
Sterilizer ID: V2178198  
03/07/2024 01:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.9 F 0.0  
2:00 166.5 F 2.8  
4:00 183.1 F 6.8  
6:00 204.3 F 11.6  
8:00 221.5 F 4.9  
10:00 229.6 F 9.7  
12:00 235.5 F 8.8  
14:00 241.6 F 11.8  
16:00 239.7 F 10.0  
18:00 252.6 F 16.4  
20:00 264.0 F 22.9  
22:00 271.1 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.6 F 28.7  
2:00 272.5 F 28.9  
3:00 271.7 F 28.8  
4:00 271.5 F 28.4

Min 271.2 F 28.4  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
5:00 195.0 F 0.0  
10:00 194.8 F 0.0  
15:00 195.7 F 0.0  
20:00 193.7 F 0.0  
25:00 191.4 F 0.0  
30:00 190.0 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 22:40  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:01:31  
CYCLE COMPLETE

03/07/2024 02:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_