

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1226

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2021 07:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
2:00 188.6 F 1.8  
4:00 191.1 F 5.1  
6:00 212.3 F 10.3  
8:00 227.5 F 5.8  
10:00 233.6 F 10.3  
12:00 237.6 F 7.9  
14:00 243.4 F 13.0  
16:00 241.8 F 11.4  
18:00 254.6 F 17.9  
20:00 265.7 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.2  
2:00 271.4 F 28.5  
3:00 272.1 F 29.3  
4:00 272.1 F 29.0  
  
Min 271.2 F 28.4  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.0 F 0.0  
5:00 233.4 F 0.0  
10:00 222.9 F 0.0  
15:00 229.2 F 0.0  
20:00 230.3 F 0.0  
25:00 229.0 F 0.0  
30:00 195.9 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 21:46  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:00:25  
CYCLE COMPLETE

10/04/2021 08:52 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01