
Midmark M11 - v1.0.5
Total Cycles: 1282

Name: RITTER2
Sterilizer ID: V2313228
07/11/2022 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.8 F 0.0
2:00 166.9 F 3.2
4:00 180.1 F 7.0
6:00 198.6 F 11.6
8:00 214.4 F 5.8
10:00 228.4 F 10.9
12:00 229.3 F 7.3
14:00 242.2 F 13.2
16:00 242.2 F 12.0
18:00 254.6 F 18.4
20:00 265.6 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.9
2:00 271.9 F 29.3
3:00 271.5 F 29.6
4:00 272.8 F 29.9

Min 271.1 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 213.4 F 0.0
10:00 230.4 F 0.0
15:00 231.4 F 0.0
20:00 227.7 F 0.0
25:00 227.6 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

07/11/2022 02:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____