
Midmark M11 - v1.0.5
Total Cycles: 3465

Name: RITTER2
Sterilizer ID: V2313228
03/21/2024 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.8 F 0.0
2:00 97.0 F 0.0
4:00 118.4 F 0.2
6:00 148.8 F 2.5
8:00 171.9 F 5.7
10:00 189.0 F 9.0
12:00 203.7 F 12.6
14:00 215.9 F 5.1
16:00 226.2 F 8.8
18:00 236.3 F 12.9
20:00 236.4 F 9.9
22:00 238.0 F 9.6
24:00 247.9 F 14.4
26:00 258.9 F 20.5
28:00 269.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.5
2:00 271.9 F 28.8
3:00 272.2 F 29.5
4:00 272.7 F 29.3

Min 271.5 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 220.3 F 0.0
10:00 225.3 F 0.0
15:00 217.9 F 0.0
20:00 216.5 F 0.0
25:00 214.9 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:11
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

03/21/2024 02:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____