
Midmark M11 - v1.0.5
Total Cycles: 3321

Name: RITTER2
Sterilizer ID: V2313228
01/25/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 72.4 F 0.0
4:00 76.9 F 0.0
6:00 93.7 F 0.5
8:00 130.4 F 2.7
10:00 159.8 F 5.5
12:00 194.1 F 8.8
14:00 206.4 F 12.8
16:00 211.9 F 5.4
18:00 222.3 F 9.2
20:00 233.2 F 13.4
22:00 234.5 F 9.7
24:00 237.1 F 9.4
26:00 246.8 F 14.0
28:00 258.0 F 20.1
30:00 268.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.4 F 29.2
2:00 272.2 F 29.6
3:00 271.6 F 28.8
4:00 272.7 F 29.5

Min 271.4 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 192.9 F 0.0
10:00 195.4 F 0.0
15:00 189.0 F 0.0
20:00 187.1 F 0.0
25:00 188.0 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:15
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

01/25/2024 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____