
Midmark M11 - v1.0.5
Total Cycles: 4076

Name: RITTER2
Sterilizer ID: V2313228
11/01/2024 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 112.5 F 0.0
4:00 125.4 F 0.1
6:00 165.9 F 2.0
8:00 192.8 F 5.8
10:00 201.1 F 9.6
12:00 216.4 F 11.8
14:00 225.2 F 7.8
16:00 235.2 F 12.2
18:00 235.4 F 9.5
20:00 238.7 F 10.1
22:00 246.6 F 14.0
24:00 257.9 F 20.1
26:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.7 F 29.7
2:00 271.4 F 28.9
3:00 272.0 F 29.7
4:00 272.3 F 29.4

Min 271.0 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 214.6 F 0.0
10:00 194.9 F 0.0
15:00 194.2 F 0.0
20:00 214.3 F 0.0
25:00 210.0 F 0.0
30:00 213.1 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 27:20
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:05
CYCLE COMPLETE

11/01/2024 03:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____