
Midmark M11 - v1.0.0
Total Cycles: 4107

Name: RITTER 1
Sterilizer ID: V2178198
12/11/2023 05:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.6 F 0.0
2:00 147.5 F 3.1
4:00 171.9 F 7.4
6:00 197.3 F 12.8
8:00 215.9 F 6.3
10:00 231.0 F 11.7
12:00 232.8 F 8.2
14:00 241.8 F 11.1
16:00 245.5 F 12.9
18:00 258.6 F 19.9
20:00 270.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.7 F 28.9
2:00 271.6 F 28.2
3:00 272.4 F 28.6
4:00 271.9 F 28.5

Min 271.5 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 205.7 F 0.0
10:00 223.4 F 0.0
15:00 222.8 F 0.0
20:00 221.3 F 0.0
25:00 221.9 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

12/11/2023 06:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____