
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001808
13:07 26/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:23
STERILIZING 3:27
136.0 C 220kPa 3:27
136.4 C 224kPa 3:57
136.2 C 223kPa 4:27
136.4 C 226kPa 4:57
136.3 C 224kPa 5:27
136.4 C 222kPa 5:57
136.5 C 224kPa 6:27
136.4 C 225kPa 6:57
136.4 C 225kPa 6:57
VENTING 6:57
AIR DRYING 7:47
CYCLE COMPLETE 17:05

Digital Signature #
C955CA7E5C39A446
