
Midmark M11 - v1.0.0
Total Cycles: 2014

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2022 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.5 F 0.0
2:00 122.5 F 0.3
4:00 127.6 F 1.8
6:00 159.0 F 5.2
8:00 186.0 F 9.2
10:00 219.4 F 8.0
12:00 224.9 F 7.5
14:00 234.8 F 12.0
16:00 232.2 F 8.2
18:00 243.5 F 12.0
20:00 242.7 F 11.9
22:00 254.9 F 18.2
24:00 265.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.5 F 29.5
2:00 272.3 F 29.0
3:00 271.2 F 28.5
4:00 271.9 F 29.3

Min 271.1 F 28.5
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 205.6 F 0.0
10:00 215.2 F 0.0
15:00 220.3 F 0.0
20:00 222.1 F 0.0
25:00 223.7 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:03:55
CYCLE COMPLETE

06/09/2022 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____