
Midmark M11 - v1.0.0
Total Cycles: 5438

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2025 04:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 167.5 F 3.3
4:00 181.5 F 7.1
6:00 209.6 F 11.9
8:00 222.5 F 5.4
10:00 233.9 F 10.0
12:00 236.6 F 8.9
14:00 242.4 F 12.1
16:00 239.9 F 10.5
18:00 252.5 F 16.8
20:00 263.4 F 23.3
22:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.0 F 29.1
2:00 270.9 F 28.5
3:00 272.2 F 29.6
4:00 272.8 F 29.5

Min 270.9 F 28.5
Max 273.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 249.0 F 0.0
10:00 230.0 F 0.0
15:00 228.6 F 0.0
20:00 224.5 F 0.0
25:00 225.7 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

03/24/2025 05:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____