
Midmark M11 - v1.0.5
Total Cycles: 3378

Name: RITTER2
Sterilizer ID: V2313228
02/12/2024 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.3 F 0.0
2:00 110.0 F 0.0
4:00 131.5 F 1.2
6:00 162.0 F 4.6
8:00 183.2 F 8.6
10:00 203.6 F 13.0
12:00 223.9 F 6.3
14:00 232.0 F 10.8
16:00 230.5 F 7.3
18:00 244.6 F 13.5
20:00 243.7 F 12.4
22:00 256.4 F 19.1
24:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.2 F 29.2
2:00 272.7 F 29.7
3:00 271.6 F 29.1
4:00 272.4 F 29.3

Min 271.3 F 28.9
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 223.0 F 0.0
10:00 205.8 F 0.0
15:00 186.6 F 0.0
20:00 178.5 F 0.0
25:00 176.7 F 0.0
30:00 176.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

02/12/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____