
Midmark M11 - v1.0.5
Total Cycles: 2794

Name: RITTER2
Sterilizer ID: V2313228
08/21/2023 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 176.6 F 0.5
4:00 187.5 F 3.8
6:00 205.2 F 8.0
8:00 220.7 F 12.8
10:00 230.5 F 8.4
12:00 240.8 F 12.4
14:00 242.0 F 11.9
16:00 241.1 F 10.9
18:00 253.5 F 17.3
20:00 264.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.0
2:00 272.9 F 29.4
3:00 271.8 F 29.3
4:00 272.0 F 29.0

Min 271.5 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 235.5 F 0.0
10:00 232.9 F 0.0
15:00 189.4 F 0.0
20:00 182.3 F 0.0
25:00 189.2 F 0.0
30:00 185.8 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

08/21/2023 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____