
Midmark M11 - v1.0.5
Total Cycles: 771

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 167.8 F 3.8
4:00 190.3 F 8.3
6:00 211.1 F 10.9
8:00 224.3 F 8.7
10:00 234.5 F 10.2
12:00 241.7 F 12.6
14:00 242.3 F 11.9
16:00 255.7 F 18.9
18:00 267.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.4 F 29.8
2:00 271.9 F 29.1
3:00 271.9 F 29.6
4:00 272.4 F 29.6

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 227.0 F 0.0
10:00 202.5 F 0.0
Dry Cycle Aborted

FILLING: 1:28
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:11
DRYING: 11:19
TOTAL CYCLE: 00:39:08

02/08/2022 02:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____