
Midmark M11 - v1.0.5
Total Cycles: 1362

Name: RITTER2
Sterilizer ID: V2313228
08/03/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 174.2 F 0.0
4:00 175.8 F 0.9
6:00 198.6 F 5.4
8:00 223.8 F 10.1
10:00 228.9 F 7.2
12:00 238.7 F 11.8
14:00 237.3 F 9.8
16:00 237.2 F 9.0
18:00 249.9 F 15.6
20:00 261.7 F 22.4
22:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.5
2:00 271.5 F 28.7
3:00 272.3 F 29.6
4:00 272.0 F 29.0

Min 271.2 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 229.0 F 0.0
10:00 237.3 F 0.0
15:00 229.1 F 0.0
20:00 225.9 F 0.0
25:00 224.9 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

08/03/2022 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____