

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000842  
16:46 05/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:46  
STERILIZING 3:54  
136.0 C 220kPa 3:54  
136.1 C 231kPa 4:24  
136.2 C 232kPa 4:54  
136.0 C 230kPa 5:24  
136.4 C 233kPa 5:54  
136.4 C 231kPa 6:24  
136.2 C 231kPa 6:54  
136.3 C 233kPa 7:24  
136.3 C 233kPa 7:24  
VENTING 7:24  
AIR DRYING 8:16  
CYCLE COMPLETE 23:16  
-----

Digital Signature #  
532531FDF5237C34  
-----