
Midmark M11 - v1.0.0
Total Cycles: 2424

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2022 05:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 175.2 F 3.9
4:00 195.2 F 8.3
6:00 213.0 F 11.0
8:00 224.3 F 7.4
10:00 236.7 F 12.4
12:00 237.4 F 9.6
14:00 237.6 F 9.0
16:00 249.1 F 14.6
18:00 261.0 F 21.3
20:00 271.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 271.4 F 28.0
3:00 272.8 F 28.9
4:00 271.9 F 28.2

Min 271.3 F 27.9
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 207.3 F 0.0
10:00 201.5 F 0.0
15:00 200.3 F 0.0
20:00 204.4 F 0.0
25:00 224.1 F 0.0
30:00 245.1 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

10/12/2022 06:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____