

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001668  
12:31 03/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:49  
STERILIZING 4:00  
136.4 C 226kPa 4:00  
136.1 C 223kPa 4:30  
136.4 C 225kPa 5:00  
136.4 C 225kPa 5:30  
136.4 C 225kPa 6:00  
136.4 C 225kPa 6:30  
136.4 C 225kPa 7:00  
136.4 C 224kPa 7:30  
136.4 C 224kPa 7:30  
VENTING 7:30  
AIR DRYING 8:20  
CYCLE COMPLETE 8:27

-----  
Digital Signature #  
FE1DF5F95DE6FE67  
-----