
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001174
13:30 11/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:15
STERILIZING 4:46
136.5 C 224kPa 4:46
136.4 C 224kPa 5:16
136.4 C 225kPa 5:46
136.4 C 224kPa 6:16
136.4 C 222kPa 6:46
136.4 C 225kPa 7:16
136.4 C 225kPa 7:46
136.1 C 222kPa 8:16
136.1 C 222kPa 8:16
VENTING 8:16
AIR DRYING 9:18
CYCLE COMPLETE 9:40

Digital Signature #
A6BACFC116AE86D3
