

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000102  
12:01 02/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:52  
PRESSURIZING 3:52  
STERILIZING 5:17  
136.0 C 223kPa 5:17  
136.4 C 227kPa 5:47  
136.2 C 224kPa 6:17  
136.4 C 224kPa 6:47  
136.4 C 225kPa 7:17  
136.3 C 225kPa 7:47  
136.4 C 227kPa 8:17  
136.4 C 227kPa 8:47  
136.4 C 227kPa 8:47  
VENTING 8:47  
AIR DRYING 9:46  
CYCLE COMPLETE 24:46  
-----

Digital Signature #  
1C8257A475CA00DC  
-----