
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005392
11:26 20/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:28
STERILIZING 3:32
136.4 C 220kPa 3:32
136.3 C 221kPa 4:02
136.4 C 220kPa 4:32
136.4 C 224kPa 5:02
136.4 C 224kPa 5:32
136.3 C 224kPa 6:02
136.4 C 221kPa 6:32
136.4 C 224kPa 7:02
136.4 C 224kPa 7:02
VENTING 7:02
AIR DRYING 7:55
CYCLE COMPLETE 8:03

Digital Signature #
90EBDE03357C26F5
