
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001612
14:37 11/24/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 5:35
STERILIZING 7:10
136.0 C 220kPa 7:10
136.4 C 217kPa 7:40
136.2 C 220kPa 8:10
136.4 C 223kPa 8:40
136.1 C 222kPa 9:10
136.4 C 223kPa 9:40
136.3 C 222kPa 10:10
136.0 C 221kPa 10:40
VENTING 10:40
AIR DRYING 11:47
CYCLE COMPLETE 19:43

Digital Signature #
F67B6542E7E70D92
