
Midmark M11 - v1.0.5
Total Cycles: 1725

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.3 F 0.0
2:00 177.6 F 2.0
4:00 193.7 F 7.4
6:00 217.0 F 13.1
8:00 229.6 F 8.4
10:00 239.8 F 10.9
12:00 245.1 F 13.4
14:00 245.9 F 13.3
16:00 259.7 F 20.8
18:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.7 F 29.4
2:00 272.0 F 28.6
3:00 272.6 F 29.4
4:00 272.1 F 28.9

Min 271.6 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 243.9 F 0.0
10:00 216.3 F 0.0
15:00 217.2 F 0.0
20:00 224.6 F 0.0
25:00 220.9 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

11/15/2022 12:02 PM
ACCEPTED: X
REJECTED:
Operator: U02