
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000427
15:48 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:29
PRESSURIZING 4:29
STERILIZING 6:01
136.4 C 223kPa 6:01
136.3 C 224kPa 6:31
136.4 C 225kPa 7:01
136.4 C 224kPa 7:31
136.2 C 223kPa 8:01
136.4 C 225kPa 8:31
136.4 C 224kPa 9:01
136.4 C 225kPa 9:31
136.4 C 225kPa 9:31
VENTING 9:31
AIR DRYING 10:38
CYCLE COMPLETE 10:41

Digital Signature #
09DDAED89BFE5F26
