
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002617
8:41 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:33
STERILIZING 3:45
136.0 C 220kPa 3:45
136.4 C 224kPa 4:15
136.4 C 224kPa 4:45
136.4 C 224kPa 5:15
136.4 C 224kPa 5:45
136.3 C 223kPa 6:15
136.3 C 223kPa 6:45
136.4 C 224kPa 7:15
136.4 C 224kPa 7:15
VENTING 7:15
AIR DRYING 8:04
CYCLE COMPLETE 10:53

Digital Signature #
32C7EA8C595B9185
