
Midmark M11 - v1.0.0
Total Cycles: 3375

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 199.4 F 4.3
4:00 217.0 F 8.9
6:00 229.1 F 10.2
8:00 231.6 F 8.5
10:00 238.7 F 11.8
12:00 240.0 F 11.1
14:00 237.8 F 9.4
16:00 251.5 F 16.1
18:00 263.1 F 22.9
20:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 28.9
1:00 272.2 F 28.8
2:00 272.2 F 28.8
3:00 272.1 F 28.9
4:00 272.2 F 29.0

Min 271.9 F 28.8
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 227.1 F 0.0
10:00 229.8 F 0.0
15:00 235.1 F 0.0
20:00 218.1 F 0.0
25:00 218.4 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

06/12/2023 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____