

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002604  
10:04 25/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:09  
STERILIZING 4:14  
136.0 C 220kPa 4:14  
136.4 C 223kPa 4:44  
136.4 C 225kPa 5:14  
136.3 C 225kPa 5:44  
136.4 C 225kPa 6:14  
136.4 C 223kPa 6:44  
136.3 C 224kPa 7:14  
136.3 C 223kPa 7:44  
136.3 C 223kPa 7:44  
VENTING 7:44  
AIR DRYING 8:35  
CYCLE COMPLETE 10:30  
-----

Digital Signature #  
E2A6D8CAB3F414B0  
-----