
Midmark M11 - v1.0.0
Total Cycles: 1246

Name: RITTER 1
Sterilizer ID: V2178198
10/08/2021 06:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 74.1 F 0.0
4:00 87.7 F 0.0
6:00 124.2 F 0.0
8:00 158.0 F 0.1
10:00 207.5 F 0.2
12:00 211.8 F 0.5
14:00 223.5 F 3.9
16:00 228.9 F 6.4
18:00 238.0 F 9.5
20:00 246.2 F 13.3
22:00 240.3 F 10.2
24:00 237.1 F 8.6
26:00 244.5 F 11.7
28:00 246.5 F 13.0
30:00 258.5 F 19.4
32:00 269.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.8
1:00 272.6 F 28.4
2:00 271.5 F 27.6
3:00 272.7 F 28.6
4:00 272.0 F 27.9

Min 271.4 F 27.6
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 222.6 F 0.0
10:00 181.2 F 0.0
15:00 183.2 F 0.0
20:00 187.5 F 0.0
25:00 233.5 F 0.0
30:00 233.1 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 33:08
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:11:08
CYCLE COMPLETE

10/08/2021 08:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____