
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001277
14:59 28/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:20
PRESSURIZING 4:21
STERILIZING 5:26
136.0 C 221kPa 5:26
136.5 C 224kPa 5:56
136.2 C 223kPa 6:26
136.4 C 225kPa 6:56
136.4 C 224kPa 7:26
136.3 C 224kPa 7:56
136.3 C 224kPa 8:26
136.4 C 224kPa 8:56
136.4 C 224kPa 8:56
VENTING 8:56
AIR DRYING 9:55
CYCLE COMPLETE 9:57

Digital Signature #
0828EEBBD9C9E4E1
