

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4171

Name: RITTER 1  
Sterilizer ID: V2178198  
12/29/2023 01:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.3 F 0.0  
2:00 172.1 F 2.6  
4:00 183.7 F 6.5  
6:00 205.3 F 12.6  
8:00 222.6 F 6.4  
10:00 237.0 F 11.8  
12:00 235.5 F 8.9  
14:00 238.9 F 9.6  
16:00 247.8 F 14.1  
18:00 260.8 F 21.3  
20:00 272.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.4 F 28.6  
2:00 272.7 F 29.1  
3:00 271.8 F 28.1  
4:00 272.9 F 29.0

Min 271.5 F 28.1  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 242.8 F 0.0  
10:00 227.2 F 0.0  
15:00 223.9 F 0.0  
20:00 223.7 F 0.0  
25:00 220.9 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 20:26  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:27  
CYCLE COMPLETE

12/29/2023 02:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_