
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002637
9:09 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:09
STERILIZING 3:12
136.0 C 221kPa 3:12
135.9 C 223kPa 3:42
136.1 C 223kPa 4:12
136.4 C 226kPa 4:42
136.4 C 225kPa 5:12
136.4 C 222kPa 5:42
136.4 C 224kPa 6:12
136.4 C 223kPa 6:42
136.4 C 223kPa 6:42
VENTING 6:42
AIR DRYING 7:33
CYCLE COMPLETE 15:01

Digital Signature #
DDCEC2598694969E
