
Midmark M11 - v1.0.0
Total Cycles: 3569

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2023 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.4 F 0.0
2:00 156.2 F 0.0
4:00 172.5 F 2.2
6:00 207.9 F 5.4
8:00 207.9 F 9.3
10:00 220.1 F 13.1
12:00 225.0 F 6.7
14:00 235.6 F 10.9
16:00 235.3 F 7.8
18:00 244.5 F 12.9
20:00 242.7 F 11.5
22:00 253.9 F 17.2
24:00 264.0 F 23.2
26:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.9
2:00 271.9 F 28.4
3:00 271.3 F 28.2
4:00 272.4 F 29.0

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 216.6 F 0.0
10:00 217.8 F 0.0
15:00 219.5 F 0.0
20:00 223.3 F 0.0
25:00 226.3 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:58
CYCLE COMPLETE

07/24/2023 01:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____