

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5088

Name: RITTER 1  
Sterilizer ID: V2178198  
11/13/2024 06:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
2:00 185.7 F 0.9  
4:00 206.0 F 5.7  
6:00 224.0 F 11.4  
8:00 229.1 F 7.0  
10:00 242.5 F 12.8  
12:00 241.3 F 10.8  
14:00 239.1 F 9.5  
16:00 254.3 F 17.0  
18:00 266.9 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 27.8  
1:00 272.0 F 27.9  
2:00 272.7 F 28.4  
3:00 272.2 F 28.3  
4:00 271.9 F 27.8  
  
Min 271.6 F 27.6  
Max 272.9 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 207.7 F 0.0  
10:00 202.1 F 0.0  
15:00 206.4 F 0.0  
20:00 206.4 F 0.0  
25:00 230.9 F 0.0  
30:00 232.8 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 19:21  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

11/13/2024 07:23 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_