

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
2.7uS / 1.7ppm  
CYCLE NUMBER 001393  
13:06 09/01/2021  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 2:48  
PRESSURIZING 10:42  
STERILIZING 11:56  
136.4 C 221kPa 11:56  
136.4 C 222kPa 12:26  
136.2 C 221kPa 12:56  
136.3 C 221kPa 13:26  
136.0 C 219kPa 13:56  
136.5 C 223kPa 14:26  
136.5 C 223kPa 14:57  
135.9 C 219kPa 15:27  
136.2 C 219kPa 15:57  
136.1 C 219kPa 16:27  
136.4 C 223kPa 16:57  
136.4 C 223kPa 17:27  
136.4 C 222kPa 17:56  
VENTING 17:56  
AIR DRYING 18:48  
CYCLE COMPLETE 18:49  
-----

Digital Signature #  
CB83F9A304A9F8ED  
-----