
Midmark M11 - v1.0.0
Total Cycles: 2839

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2023 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 183.7 F 3.2
4:00 195.7 F 8.2
6:00 214.1 F 13.4
8:00 223.7 F 7.9
10:00 239.3 F 13.1
12:00 238.2 F 10.4
14:00 235.8 F 8.4
16:00 250.3 F 15.4
18:00 262.2 F 22.2
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.1
2:00 271.8 F 28.3
3:00 272.5 F 29.2
4:00 272.4 F 28.8

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 230.3 F 0.0
10:00 239.3 F 0.0
15:00 233.1 F 0.0
20:00 227.5 F 0.0
25:00 230.8 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

02/01/2023 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____