

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001471  
10:41 25/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:46  
STERILIZING 3:55  
136.0 C 220kPa 3:55  
136.4 C 225kPa 4:25  
136.4 C 226kPa 4:55  
136.4 C 226kPa 5:25  
136.4 C 225kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 222kPa 6:55  
136.4 C 225kPa 7:25  
136.4 C 225kPa 7:25  
VENTING 7:25  
AIR DRYING 8:20  
CYCLE COMPLETE 8:40  
-----

Digital Signature #  
618E8EACD8E3C791  
-----