

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4378

Name: RITTER2  
Sterilizer ID: V2313228  
02/25/2025 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.9 F 0.0  
2:00 123.6 F 1.1  
4:00 123.4 F 3.5  
6:00 144.1 F 7.2  
8:00 173.3 F 11.5  
10:00 204.3 F 4.1  
12:00 225.8 F 8.4  
14:00 230.8 F 12.9  
16:00 232.5 F 9.7  
18:00 236.9 F 9.5  
20:00 244.7 F 13.8  
22:00 256.3 F 20.0  
24:00 267.2 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 30.1  
1:00 272.1 F 30.3  
2:00 272.0 F 30.1  
3:00 272.1 F 30.3  
4:00 272.1 F 30.1

Min 271.6 F 29.9  
Max 272.2 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.2 F 0.0  
5:00 206.9 F 0.0  
10:00 202.1 F 0.0  
15:00 192.3 F 0.0  
20:00 183.4 F 0.0  
25:00 173.8 F 0.0  
30:00 172.3 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:04:06  
CYCLE COMPLETE

02/25/2025 01:48 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_