

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4030

Name: RITTER2  
Sterilizer ID: V2313228  
10/16/2024 08:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 71.3 F 0.0  
4:00 92.4 F 0.0  
6:00 138.1 F 0.1  
8:00 146.1 F 0.4  
10:00 169.3 F 1.1  
12:00 186.5 F 3.7  
14:00 204.6 F 6.5  
16:00 215.2 F 10.0  
18:00 224.3 F 12.4  
20:00 227.0 F 8.2  
22:00 236.5 F 12.4  
24:00 235.4 F 9.5  
26:00 239.6 F 10.2  
28:00 246.6 F 13.8  
30:00 257.7 F 19.8  
32:00 267.8 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.4 F 29.5  
2:00 271.6 F 28.8  
3:00 271.5 F 29.2  
4:00 272.3 F 29.4  
  
Min 271.1 F 28.6  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 188.9 F 0.0  
10:00 199.4 F 0.0  
15:00 190.0 F 0.0  
20:00 194.4 F 0.0  
25:00 197.6 F 0.0  
30:00 196.3 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 33:27  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:11:55  
CYCLE COMPLETE

10/16/2024 10:08 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_