
Midmark M11 - v1.0.5
Total Cycles: 508

Name: RITTER2
Sterilizer ID: V2313228
11/01/2021 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.0 F 0.0
2:00 140.2 F 0.0
4:00 141.4 F 1.6
6:00 158.9 F 5.7
8:00 188.2 F 10.7
10:00 211.6 F 5.0
12:00 223.3 F 10.6
14:00 230.3 F 7.4
16:00 238.6 F 12.0
18:00 242.9 F 12.7
20:00 255.3 F 19.2
22:00 266.2 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.6
1:00 272.3 F 30.1
2:00 271.3 F 29.6
3:00 272.1 F 30.2
4:00 271.6 F 29.5

Min 271.3 F 29.5
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 240.4 F 0.0
10:00 235.1 F 0.0
15:00 218.7 F 0.0
20:00 202.6 F 0.0
25:00 194.3 F 0.0
30:00 187.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:34
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

11/01/2021 09:38 AM
ACCEPTED: X
REJECTED:
Operator: U01