

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1847

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2022 12:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.6 F 0.0  
2:00 175.5 F 3.2  
4:00 187.6 F 8.2  
6:00 209.7 F 9.4  
8:00 226.8 F 9.0  
10:00 234.6 F 9.7  
12:00 242.6 F 12.8  
14:00 243.0 F 12.1  
16:00 256.4 F 19.2  
18:00 268.3 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.7 F 29.5  
2:00 271.5 F 28.9  
3:00 272.9 F 29.7  
4:00 272.1 F 29.1  
  
Min 271.5 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 206.8 F 0.0  
10:00 220.9 F 0.0  
15:00 227.8 F 0.0  
20:00 229.6 F 0.0  
25:00 228.2 F 0.0  
30:00 222.9 F 0.0  
  
Drying Complete

FILLING: 2:13  
HEATING: 19:11  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

12/15/2022 01:07 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_