

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5028

Name: RITTER 1  
Sterilizer ID: V2178198  
10/28/2024 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 69.5 F 0.0  
4:00 82.3 F 0.0  
6:00 121.1 F 0.1  
8:00 149.4 F 0.4  
10:00 171.3 F 1.2  
12:00 188.5 F 3.6  
14:00 201.5 F 6.5  
16:00 209.2 F 9.7  
18:00 220.6 F 13.2  
20:00 225.1 F 6.8  
22:00 235.2 F 10.9  
24:00 233.2 F 7.4  
26:00 244.6 F 12.7  
28:00 241.8 F 10.9  
30:00 253.8 F 16.9  
32:00 264.2 F 23.0  
34:00 271.9 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.6 F 28.6  
2:00 271.8 F 27.9  
3:00 272.0 F 28.4  
4:00 272.4 F 28.4

Min 271.4 F 27.7  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.3 F 0.0  
5:00 187.9 F 0.0  
10:00 182.2 F 0.0  
15:00 196.7 F 0.0  
20:00 208.8 F 0.0  
25:00 219.3 F 0.0  
30:00 224.2 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 34:09  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:12:30  
CYCLE COMPLETE

10/28/2024 10:07 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_