
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002872
16:14 16/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:20
STERILIZING 3:21
136.4 C 220kPa 3:21
136.3 C 224kPa 3:51
136.3 C 224kPa 4:21
136.3 C 223kPa 4:51
136.4 C 225kPa 5:21
136.4 C 223kPa 5:51
136.5 C 226kPa 6:21
136.4 C 226kPa 6:51
136.4 C 226kPa 6:51
VENTING 6:52
AIR DRYING 7:44
CYCLE COMPLETE 7:52

Digital Signature #
12DDD7FA286FAAB6
