
Midmark M11 - v1.0.5
Total Cycles: 3024

Name: RITTER2
Sterilizer ID: V2313228
10/18/2023 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.5 F 0.0
2:00 132.5 F 1.0
4:00 142.9 F 3.6
6:00 169.5 F 7.5
8:00 192.6 F 12.2
10:00 212.7 F 5.4
12:00 225.9 F 10.0
14:00 230.8 F 8.5
16:00 240.4 F 12.2
18:00 239.9 F 10.8
20:00 253.3 F 17.6
22:00 264.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 30.0
2:00 272.4 F 29.4
3:00 271.3 F 28.9
4:00 271.8 F 29.7

Min 271.2 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 209.3 F 0.0
10:00 214.6 F 0.0
15:00 216.5 F 0.0
20:00 212.5 F 0.0
25:00 212.0 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:54
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:03
CYCLE COMPLETE

10/18/2023 10:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____