
Midmark M11 - v1.0.0
Total Cycles: 5266

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2025 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 175.9 F 2.9
4:00 180.2 F 7.1
6:00 200.7 F 11.5
8:00 212.5 F 4.2
10:00 225.7 F 8.9
12:00 238.0 F 12.5
14:00 237.4 F 10.2
16:00 236.9 F 8.6
18:00 247.4 F 14.1
20:00 258.7 F 20.2
22:00 269.0 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.2
2:00 272.2 F 28.8
3:00 271.4 F 28.3
4:00 272.0 F 29.1

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 204.1 F 0.0
10:00 201.3 F 0.0
15:00 198.6 F 0.0
20:00 202.2 F 0.0
25:00 211.4 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

01/23/2025 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____