
Midmark M11 - v1.0.0
Total Cycles: 4818

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2024 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 179.4 F 2.7
4:00 203.0 F 6.6
6:00 212.4 F 11.2
8:00 222.5 F 4.8
10:00 230.6 F 9.5
12:00 239.5 F 9.7
14:00 240.7 F 11.7
16:00 239.0 F 10.0
18:00 252.1 F 16.4
20:00 263.2 F 22.8
22:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.8 F 29.4
2:00 272.2 F 28.8
3:00 272.1 F 29.1
4:00 272.7 F 29.3

Min 271.7 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 230.7 F 0.0
10:00 225.4 F 0.0
15:00 234.3 F 0.0
20:00 228.9 F 0.0
25:00 231.6 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

08/14/2024 04:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____