
Midmark M11 - v1.0.5
Total Cycles: 1231

Name: RITTER2
Sterilizer ID: V2313228
06/28/2022 06:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.0 F 0.0
2:00 85.1 F 0.0
4:00 106.6 F 0.0
6:00 127.1 F 0.0
8:00 160.8 F 0.2
10:00 181.3 F 1.5
12:00 198.9 F 4.9
14:00 218.3 F 8.3
16:00 227.7 F 11.9
18:00 227.4 F 6.9
20:00 238.3 F 11.4
22:00 234.6 F 8.5
24:00 242.7 F 11.5
26:00 245.2 F 12.9
28:00 256.3 F 18.9
30:00 266.5 F 25.2
32:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.7 F 29.3
2:00 271.7 F 29.0
3:00 272.5 F 29.2
4:00 271.8 F 29.1

Min 271.5 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 241.1 F 0.0
10:00 218.0 F 0.0
15:00 215.0 F 0.0
20:00 204.0 F 0.0
25:00 203.2 F 0.0
30:00 188.8 F 0.0

Drying Complete

FILLING: 1:02
HEATING: 32:28
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:10:36
CYCLE COMPLETE

06/28/2022 07:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____