

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3125

Name: RITTER 1  
Sterilizer ID: V2178198  
04/19/2023 07:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.6 F 0.0  
2:00 86.8 F 0.0  
4:00 99.0 F 0.0  
6:00 133.6 F 0.0  
8:00 180.3 F 0.0  
10:00 201.8 F 1.6  
12:00 213.0 F 5.0  
14:00 217.6 F 8.1  
16:00 227.2 F 11.5  
18:00 228.3 F 6.1  
20:00 239.6 F 11.0  
22:00 232.6 F 7.3  
24:00 246.6 F 13.4  
26:00 243.6 F 11.9  
28:00 256.0 F 18.3  
30:00 267.1 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.0  
1:00 272.4 F 28.8  
2:00 271.2 F 27.9  
3:00 272.4 F 28.7  
4:00 271.4 F 27.9

Min 271.0 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.6 F 0.0  
5:00 216.8 F 0.0  
10:00 219.6 F 0.0  
15:00 226.9 F 0.0  
20:00 206.9 F 0.0  
25:00 203.5 F 0.0  
30:00 201.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:18  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:09:21  
CYCLE COMPLETE

04/19/2023 08:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_