
Midmark M11 - v1.0.5
Total Cycles: 2218

Name: RITTER2
Sterilizer ID: V2313228
04/04/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.6 F 0.0
2:00 165.8 F 3.3
4:00 183.6 F 7.6
6:00 206.0 F 12.9
8:00 223.9 F 7.8
10:00 237.3 F 13.4
12:00 238.4 F 11.3
14:00 238.7 F 10.2
16:00 252.9 F 17.2
18:00 264.8 F 24.4
20:00 270.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 271.8 F 29.6
2:00 273.0 F 30.1
3:00 272.7 F 29.7
4:00 272.0 F 29.4

Min 271.4 F 29.1
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.2 F 0.0
5:00 239.2 F 0.0
10:00 220.2 F 0.0
15:00 193.7 F 0.0
20:00 191.5 F 0.0
25:00 190.7 F 0.0
30:00 189.6 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

04/04/2023 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____